
Testing Lab Analyzer

Designed to address the service requirements of analyzers used as primary calibrating tool

Calibrate weekly
Full cleaning every 5-6 months
Replace O2 sensor every 2 years or as necessary
Replace CO, NO, NO2 sensor every 3 years or as necessary
Replace O-rings and internal tubing of unit as necessary
Replace filters as necessary

Light to Medium Duty Service

Designed to address the service requirements of occasional usage for the flue gas analyzer.

Calibrate every 5-6 months
Full cleaning every 5-6 months
Replace O2 sensor every 2 years or as necessary
Replace CO, NO, NO2 sensor every 3 years or as necessary
Replace O-rings and internal tubing of unit as necessary
Replace filters as necessary

Medium to Heavy Duty Service

Designed to address the service requirements of consistent usage and oil fired burners with the flue gas analyzer.

Calibrate every 2-3 months
Full cleaning every 6 months
Replace O2 sensor every 2 years or as necessary
Replace CO, NO, NO2 sensor every 3 years or as necessary
Replace O-rings and internal tubing of unit as necessary
Replace filters as necessary

Heavy Duty to Severe Service

Designed to address the service requirement of a well used flue gas analyzer

Calibrate every month
Full cleaning every service
Replace O2 sensor every 2 years or as necessary
Replace CO, NO, NO2 sensor every 3 years or as necessary
Replace O-rings and internal tubing of unit as necessary
Replace filters as necessary

K-W Electronic Service Inc.

Work Order

Date: _____

750 McMurray Rd. Waterloo, Ontario, Canada, N2V 2G5 Tel: (519) 747-5042 Fax: (519) 747-4448, info@kwelectronicsservice.com

CUSTOMER INFORMATION

Name _____
Address _____
City _____
Province/State _____
Postal/ZIP _____
Contact Name _____
Phone Number _____

ANALYSER INFORMATION

Model Number _____
Serial Number _____
Previous Service Date _____
Next Service Due Date _____

Note:

Requested Service

Work Performed

Description	Time

Materials Used

KWE P/N	Description	Qty.